

Uricult Model Chart

Simple on-site urine culturing



AIDIAN

Uricult

Culture based system for detection of bacteriuria

- Ordering information: 67404 Uricult, 10 tests
- CLED medium for total bacterial counts
- Selective MacConkey medium for gram negative bacteria
- Detection range 10^3 – 10^7 CFU/ml
- Storage at 7–25°C avoiding draughts and temperature fluctuations
- Shelf life of nine months from the date of manufacture
- Incubation at $36 \pm 2^\circ\text{C}$ for 16–24 hours, additional 24 hour incubation may be required in some cases

Sample collection

Urine for bacterial culture should remain in the bladder for four hours prior to sampling. Inoculate the sample immediately after collection or refrigerate (2–8°C) the sample for up to 24 hours prior to inoculation. Some medication (especially antibiotics) may affect the results. Therefore, the sample should not be collected until 48 hours after the final dose of medication.

For more detailed information, please consult the instructions for use.

It's so easy to use!



Dip or Pour

Incubate

Interpret

AIDIAN

Aidian and Uricult are trademarks or registered trademarks of Aidian Oy. Aidian develops, manufactures, and markets diagnostic test systems for healthcare professionals with over 50 years of experience.


The products are CE-marked either under IVDD (EC/98/79) or IVDR (2017/746).

aidian.eu

Uricult

Total bacterial count should be determined on CLED medium. For further information of interpretation of results refer to the Instructions for use.

Colony densities *Escherichia coli* CFU/ml



Organisms on media The images are for illustrative purposes and show examples of how microbial growth may appear on Uricult agars.

